



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

A.K.O., Inc.

50 Baker Hollow Rd, Windsor, CT 06095

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

**ISO/IEC 17025:2017
& Meets the Requirements of ANSI/NCSL Z540.1-1994
& ANSI/NCSL Z540.3-2006**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Mechanical, Mass, Force & Weighing Devices, and Dimensional Calibration
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

February 4, 2015

Issue Date:

December 16, 2022

Expiration Date:

January 31, 2025

Accreditation No.:

79689

Certificate No.:

L22-861

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjlabs.com



Certificate of Accreditation: Supplement

A.K.O., Inc.

50 Baker Hollow Rd, Windsor, CT 06095
 Contact Name: Julie Wiezalis Phone: 860-298-9765

Accreditation is granted to the facility to perform the following calibrations:

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Pressure Transducers ^F	20 psi to 10 000 psi	0.03 % of reading	AMETEK Precision Dead Weight Testers ⁹ with Certified Weights WI-103
Torque Drivers ^F	2 ozf•in to 240 lbf•in	1.8 % of reading	TSD Torque Master Calibration Systems ⁷ with TSD Transducers WI-102
Torque Multipliers ^F	200 lbf•ft to 150 000 lbf•ft	1.7 % of reading	TSD Torque Master Calibration Systems ⁷ with TSD Transducers WI-102
Torque Transducers ^{FO}	70 gf•cm to 23 000 gf•cm	0.013 % of reading	TSD Calibration Arms & NIST Class "F" or ASTM equivalent Weights ⁶ WI-101
	1 ozf•in to 320 ozf•in	0.013 % of reading	
	0.2 lbf•in to 20 lbf•in	0.013 % of reading	
	0.5 lbf•in to 1 500 lbf•in	0.014 % of reading	
	1 lbf•ft to 300 lbf•ft	0.02 % of reading	
	5 lbf•ft to 1 000 lbf•ft	0.02 % of reading	
	20 lbf•ft to 2 500 lbf•ft	0.018 % of reading	
	20 lbf•ft to 8 000 lbf•ft	0.012 % of reading	
	100 lbf•ft to 900 lbf•ft	0.023 % of reading	
	200 lbf•ft to 1 800 lbf•ft	0.023 % of reading	
	1 000 lbf•ft to 10 000 lbf•ft	0.012 % of reading	
	2 000 lbf•ft to 20 000 lbf•ft	0.012 % of reading	
	500 lbf•ft to 3 500 lbf•ft	0.037 % of reading	
4 000 lbf•ft to 150 000 lbf•ft	0.015 % of reading		
Torque Transducers ^F	100 lbf•ft to 100 000 lbf•ft	0.03 % of reading	AKO Inline Reaction Systems ⁶ with TSD Transducers WI-101



Certificate of Accreditation: Supplement

A.K.O., Inc.

50 Baker Hollow Rd, Windsor, CT 06095
 Contact Name: Julie Wiezalis Phone: 860-298-9765

Accreditation is granted to the facility to perform the following calibrations:

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Torque Wrenches ^F	10 lbf•in to 4 000 lbf•ft	1.2 % of reading	TSD Torque Master Calibration Systems ⁷ with TSD Transducers WI-102
Torque Wrenches – Hydraulic/Pneumatic ^F	200 lbf•ft to 20 000 lbf•ft	0.32 % of reading	TSD Torque Master Calibration Systems with TSD Pressure Transducers and TSD Torque Transducers WI-104
Torque Wrenches - Hydraulic ^F	5 000 lbf•ft to 150 000 lbf•ft	0.4 % of reading	
Torque Wrenches / Multipliers with Built-In Indicators ^F	60 lbf•ft to 1 500 lbf•ft	1 % of reading	TSD Torque Master Calibration Systems with TSD Transducers WI-102
	400 lbf•ft to 6 500 lbf•ft	0.6 % of reading	

Mass, Force and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Weights Field Service NIST Class F ^F or ASTM Class 5, 6, 7 ^F	0.2 lb to 5 lb	0.000 04 lb	Balance or Scale ⁸ with ASTM Class III Reference Weights ⁸ WI-105
	5 lb to 50 lb	0.000 17 lb	

Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Calibration (AKO Moment) Arms ^F	5 in	5.8×10^{-4} in	Micrometer Rods Dial Indicator SP-106
	10 in	5.8×10^{-4} in	
	12 in	5.8×10^{-4} in	
	24 in	6.2×10^{-4} in	
	30 in	6.2×10^{-4} in	
	60 in	6.7×10^{-4} in	
	120 in	8×10^{-4} in	



Certificate of Accreditation: Supplement

A.K.O., Inc.

50 Baker Hollow Rd, Windsor, CT 06095
Contact Name: Julie Wiezalis Phone: 860-298-9765

Accreditation is granted to the facility to perform the following calibrations:

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this calibration at its fixed location.
4. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer^{FO} would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
5. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.
6. The calibration procedures referenced for Torque Transducers are ASTM/ANSI E 2624 & E 2428.
7. The calibration procedures referenced for Torque Wrenches/Drivers/ Multipliers are ASME B107.300 and ISO 6789.
8. The calibration procedure referenced for Mass is NIST IR 6969 SOP 8.
9. The calibration procedure referenced for Pressure is ASTM/ANSI E 2624.